

## Xpert qDetect Chicken

#GDK13.0100 (100 rxns)  
 (FOR RESEARCH ONLY)



**Product:** According to the European Commission Directive 2002/86/EC, food ingredients must be declared and in the case of meat products, the specification of the animal species has to be disclosed on the label. Species authenticity can be extremely relevant to consumer for a variety of reasons, including economic, medical and religious reasons. Fraudulent substitution by less expensive meats, inclusion of meat in vegetarian products or the presence of allergens are all issues of major concern.

Several methods for species detection that are available are based on protein identification (by means of electrophoretic and/or immunological methods). However, these methods are not reliable for highly processed and heated products due to protein denaturation and degradation.

As DNA is much more stable, PCR/qPCR amplification of a species-specific target sequence provide a simple, fast and reliable method for the detection of a target species with high sensitivity and specificity, even in the case of highly fragmented DNA.

The Xpert qDetect Chicken Kit is a kit for the detection of DNA from Chicken (*Gallus gallus*) present in total DNA previously purified from food samples, based on qPCR using precisely designed specific primers and FAM-labeled Taqman® probe. This kit is compatible with instruments equipped with FAM and ROX channels. The detection limit is 10 pg of Chicken DNA, allowing the detection of as little as 0.1% of target DNA in food samples if 100ng of total DNA (mixed species) is used

**Applications:** Detection of Chicken DNA in food and other samples.

**Contents:** The qPCR Detection Kit (#GDK13.0100) for Chicken contains sufficient reagents for 100 qPCR reactions.

Component	GDK13.0100
Chicken Mix A	2x 840 µl
Chicken Mix B	210 µl
Positive Control (Chicken+)	70 µl
Negative Control (Chicken-)	70 µl

**Note:** This product does not include reagents and other materials required for DNA extraction.

**Samples:** 2µl of DNA (previously purified from food products using commercially available DNA purification kits according to manufacturer's instructions or author's instructions).

**Properties:** Fast, Easy and Reliable  
 Low limit of Detection  
 100% Specificity, determined using DNA from 13 animal and 3 vegetable species commonly found in same food products.  
 Compatible with instruments equipped with **FAM** and **ROX** channels

**Storage:** -20°C and protected from light for at least 1 year. Minimize repeated freeze/thawing, consider preparation of aliquots.

## Prior to use:

### General Considerations

In order to prevent erroneous results due to contamination and/or degradation of DNA, it is strongly recommended to have dedicated materials, equipment and area for the DNA extraction, separated from the preparation of PCR procedures. Workflow in the laboratory should be unidirectional, starting with DNA isolation and moving to preparation of PCR amplification and finally to (q)PCR and Detection. Moreover, all molecular biology procedures, including DNA extraction, require qualified personnel to prevent risk of erroneous results and/or inadequate interpretation of results. Disposable gloves must be worn during all procedures. All relevant National and Local Regulations must be met.

### Food Sample Preparation

Representative food samples should be prepared conform standard procedures and under GLP conditions.

### DNA extraction

DNA is extracted using an appropriate kit (not included) for the extraction of DNA from food samples using the manufacturer's instructions and/or according to literature following author's instructions.

### Signal Detection

Unambiguous detection of Chicken DNA (and of the internal control) requires a suitable calibration of both FAM and ROX channel. Please refer to the manufacturer's instruction of the real time PCR cycler.

## qDetect - Basic Protocol

### 1. Mix for each qPCR reaction:

Component	Volume
Chicken Mix A	16 µl
Chicken Mix B	2 µl

In order to minimize risk of contamination, reagent loss and improve pipetting accuracy, we recommend to prepare a mastermix for multiple samples (N), always including a negative control, and a positive control, by mixing all components (N+2), except template DNA (nor control DNA), dividing the mixture equally into each PCR tube (18 µl each), briefly spin tubes (or tap down) and then add 2 µl template DNA or control DNA directly in the mixture.

### 2. Set-up qPCR cycling:

N° cycles	Temp	Time	Acquisition
1x	50°C	2 min	No
1x	95°C	5 min	No
40x	95°C	30 sec	No
	60°C	30 sec	Yes
	72°C	30 sec	No

After an initial cycle of 2 min at 50°C and 5 min at 95°C (Enzyme activation and denaturation of template), cycle 40 times for 30 seconds at 95°C, 30 seconds at 60°C and 30 seconds at 72°C. Acquire data for the detection of Chicken DNA on the **FAM** channel. Probe to detect specific amplification of the internal control, which is included in Chicken Mix A and which is amplified simultaneously with the target DNA, should be detected in the **ROX** channel.

## Results

### Controls

In order to validate the assay, controls must have the following results. If the signal of one of the controls does not match, the whole experiment, including all samples, must be repeated.

Control	FAM channel	ROX channel
Negative Control	Ct=N/A	positive
Positive Control	positive	unimportant*

N/A = Not applicable (signal below threshold).

\*unimportant: is expected to be positive, however, if negative but other controls match expected results, this makes no difference

### Samples

For each sample, there are 4 possible outcomes, as summarized in the table below.

FAM channel	ROX channel	Result
positive	positive	positive
positive	Ct=N/A	positive
Ct=N/A	positive	negative
Ct=N/A	Ct=N/A	inhibition*

\*) in case both Chicken (FAM channel) and Internal Control (ROX channel) have signals below threshold, but all the controls resulted in signals as expected, the sample must be retested, as the qPCR reaction was inhibited. Inhibition often is the result of a too high DNA concentration and therefore it is recommended that retesting should be carried out with a 10-fold dilution of the original DNA sample.